


## **Supporting Information**

### **Determination of Percent Hemoglobin A1c Using a Potentiometric Method**

Hong Liu and Richard M. Crooks\*



**Figure S1.** CV obtained in 0.10 M N-ethylmorpholine buffer (pH 8.5) containing 1.0 mM ARS and 0.10 M KCl. The solution was deoxygenated by bubbling nitrogen for 40 min. Scan rate: 0.100 V/s.